

Quattro™ 2.5

INSTRUCTION MANUAL

CAUTION: United States Federal Law restricts this device to sale by or on the order of a physician



This manual is valid for the Quattro™ 2.5

This user manual is published by Roscoe Medical, Inc. Roscoe Medical, Inc. and reserves the right to improve and amend it at any time without prior notice. Amendments will be published in a new edition of this manual.

All Rights Reserved.Rev. V1.0 © 2013, 20131216

Declaration of conformity:

Roscoe Medical, Inc. declares that the Quattro™ 2.5 complies with the following normative documents:

IEC60601-1, IEC60601-1-2, IEC60601-2-10, ISO 7010, ISO14971,
ISO10993-1, ISO10993-5, ISO10993-10

Complies with MDD 93/42/EEC and Amended by directive 2007/47/EC requirements

Contents

1. FOREWORD	4
2. SAFETY INFORMATION	4
3. INDICATIONS FOR USE	7
4. PRESENTATION	8
5. INSTALLATION	11
6. OPERATION	12
7. MAINTENANCE.....	36
8. TROUBLESHOOTING	37
9. SPECIFICATIONS	39
10. STORAGE	41
11. DISPOSAL	42
12. EMC TABLES	42
13. WARRANTY	46
14. NORMALIZED SYMBOLS	47

1. FOREWORD

1.1 General information

The microprocessor controlled Quattro™ 2.5 provides Interferential (4- pole interferential), Premodulated (2-pole interferential), medium frequency Russian (S and A), EMS (S and A) and TENS waveforms. You can choose between several different amplitude modulation options. The Interferential and Premodulated waveforms offer frequency modulation as well as a static frequency option. There are four output channels available with the Quattro™ 2.5. Channel 1 and 2 are grouped together and are independent of the grouped channels 3 and 4. This design feature enables the Quattro™ 2.5 to be used on two patients with different waveforms, programs and output intensities simultaneously.

1.2 Introduction to This Manual

This manual has been written for the users of the Quattro™ 2.5. It contains general information on operation, precautionary practices, and maintenance. In order to maximize its use, efficiency, and the life of the stimulator, please read this manual thoroughly and become familiar with the controls, as well as the accessories before operating the stimulator.

2. SAFETY INFORMATION

2.1 CAUTION

- 1) Keep yourself informed of the contraindications.
- 2) Read, understand, and practice the warnings, cautions and operating instructions. Know the limitations and hazards associated with using any device. Observe the precautionary and operational decals placed on the unit.
- 3) DO NOT operate this unit in an environment where other devices used intentionally radiate electromagnetic energy in an unshielded manner.
- 4) DO NOT use sharp objects such as a pencil point or ballpoint pen to operate the buttons on the control panel.
- 5) Inspect applicator cables and associated connectors before each use.
- 6) This device should not be used adjacent to or stacked with other equipment. If adjacent to or stacked equipment use is necessary, this device should be observed to verify that it is operating within the normal configuration in which it is to be used.
- 7) This device needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided in the manual.
- 8) Portable and mobile RF communications equipment can affect this device. Do not use a mobile phone or other device that emits electromagnetic fields, near the unit. This may result in incorrect operation of the device.
- 9) This device has been thoroughly tested and inspected to assure proper performance and operation!

**2.2
WARNING**

- 1) United States Federal Law restricts this device to sale to, or on the order of, a physician or licensed practitioner. This device should be used only under the continued supervision of a physician or licensed practitioner.
- 2) Make certain the unit is electrically grounded by connecting only to a grounded electrical service receptacle conforming to the applicable national and local electrical codes.
- 3) Care must be taken when operating this equipment around other equipment. Potential electromagnetic or other interference could occur to this device or to the other equipment. Try to minimize this interference by not using the other equipment in conjunction with it.
- 4) Before administering any treatment to a patient you should become acquainted with the operating procedures for each mode of treatment available, as well as the indications, contraindications, warnings and precautions. Consult other resources for additional information regarding the application of electrotherapy.
- 5) To prevent electrical shock, disconnect the unit from the power source before attempting any maintenance procedures.
- 6) The use of accessories, transducers and cables other than those specified, with the exception of transducers and cables sold by the manufacturer as replacement parts for internal components, may result in increased emissions or decreased immunity of the device.
- 7) This device is not designed to be use in an MRI environment and should be removed prior to MRI exposure.
- 8) Do not apply stimulation over the patient's neck because this could cause severe muscle spasms resulting in closure of the airway, difficulty in breathing, or adverse effects on heart rhythm or blood pressure.
- 9) Do not apply stimulation across the patient's chest, because the introduction of electrical current into the chest may cause rhythm disturbances to the patient's heart, which could be lethal.
- 10) Do not apply stimulation over open wounds or rashes, or over swollen, red, infected, or inflamed areas or skin eruptions (e.g., phlebitis, thrombophlebitis, varicose veins).
- 11) Do not apply stimulation over, or in proximity to, cancerous lesions.
- 12) Do not apply stimulation in the presence of electronic monitoring equipment (e.g., cardiac monitors, ECG alarms), which may not operate properly when the electrical stimulation device is in use.
- 13) Do not apply stimulation when the patient is in the bath or shower.
- 14) Do not apply stimulation while the patient is sleeping.
- 15) Do not apply stimulation while the patient is driving, operating machinery, or during any activity in which electrical stimulation can put the patient at risk for injury.
- 16) Consult with the patient's physician before using this device, because the device may cause lethal rhythm disturbances to the heart insusceptible individuals.
- 17) Apply stimulation only to normal, intact, clean, dry, healthy skin.
- 18) This device should not be used for symptomatic local pain relief unless etiology is established or unless a pain syndrome has been diagnosed.

- 19) Patients with arterial or venous thrombosis or thrombophlebitis are at risk of developing embolisms when electrical stimulation is applied over or adjacent to the vessels containing the thrombus. If a patient has a history of deep vein thrombosis, even many years past, the affected area should not be stimulated.
- 20) Fresh fractures should not be stimulated in order to avoid unwanted motion.
- 21) Stimulation should not be applied immediately following trauma or to tissues susceptible to hemorrhage.
- 22) Do not apply electrodes directly over the eyes or inside body cavities.
- 23) Do not use electrical stimulation in conjunction with high frequency surgical equipment or microwave or shortwave therapy systems.
- 24) Keep electrodes separated during treatment. Electrodes in contact with each other could result in improper stimulation or skin burns.
- 25) Since the effects of stimulation of the brain are unknown, stimulation should not be applied across the head, and electrodes should not be placed on opposite sides of the head.

2.3 CONTRA- INDICATIONS

- 1) Do not use this device on patients who have a cardiac pacemaker, implanted defibrillator, or other implanted metallic or electronic device, because this may cause electric shock, burns, electrical interference, or death.
- 2) Do not use this device on patients whose pain syndromes are undiagnosed.

2.4 PRECAUTIONS

- 1) Federal law (USA) restricts this device to sale to or on the order of a physician.
- 2) The long-term effects of chronic electrical stimulation are unknown.
- 3) Electrical stimulation devices have no curative value.
- 4) Electrical stimulation is not a substitute for pain medications and other pain management therapies.
- 5) Effectiveness is highly dependent upon the patient and the selection of a practitioner qualified in the management of pain.
- 6) The safety of electrical stimulation during pregnancy has not been established.
- 7) Some patients may experience skin irritation or hypersensitivity due to the electrical stimulation or electrical conductive medium (gel).
- 8) Physician recommended precautions should be followed when treating patients with suspected or diagnosed heart disease.
- 9) Physician recommended precautions should be followed when treating patients with suspected or diagnosed epilepsy.
- 10) Physician recommended precautions should be followed when treating patients who have a tendency to bleed internally, such as following an injury or fracture.
- 11) Physician recommended precautions should be followed when treating patients who recently had any surgical procedures; stimulation may disrupt the patient's healing process.
- 12) Use caution if stimulation is applied over the menstruating or pregnant uterus.
- 13) Use caution if stimulation is applied over areas of skin that lacks normal sensation.
- 14) Use of this device on a patient should only be used under the continued supervision of a licensed practitioner.
- 15) Electrical stimulation is ineffective for pain of central origin.

- 16) Use extreme caution when treating desensitized areas on patients who may not be able to report discomfort or pain.
- 17) Patients should not be left unattended during any treatment.
- 18) Keep this device out of the reach of children.

**2.5
Adverse
Reaction**

- 1) Skin irritation, inflammation, and electrode burns beneath the electrodes are potential adverse reactions.
- 2) Patients may experience headache and other painful sensations during, or following the application of, electrical stimulation near the eyes and to the head and face.
- 3) The clinician should stop using this device and should consult with the patient's attending physician should the patient experience any adverse reactions from treatment used with this device.

**3.
INDICATIONS
FOR USE**

For TENS, Interferential and Premodulated(IFC):



- 1) Symptomatic relief of chronic intractable pain
- 2) Post-traumatic acute pain
- 3) Post-surgical acute pain

Additionally for EMS and Russian:

- 1) Relaxation of Muscle spasms
- 2) Prevention or retardation of disuse atrophy
- 3) Increasing local blood circulation
- 4) Muscle re-education
- 5) Maintaining or increasing range of motion
- 6) Immediate postsurgical stimulation of calf muscles to prevent venous thrombosis

4. PRESENTATION

4.1 Panel for Front View



- 1) Channel Button:
 - Press to enter treatment parameters for channel 1 and 2.
 - During treatment, press to display treatment parameters for channel 1 and 2.
 - Rotate to adjust the output intensity for channel 1.
- 2) Channel Button:
 - Press to enter treatment parameters for channel 1 and 2.
 - During treatment, press to display treatment parameters for channel 1 and 2.
 - Rotate to adjust the output intensity for channel 2.
- 3) Parameters Control Knob and Pause button.
- 4) Stop Treatment Button: Press to stop current channels' treatment.
- 5) Power Indicator Light
- 6) LCD display: Shows the current information of the device.
- 7) Channel Button:
 - Press to enter treatment parameters for channel 3 and 4.
 - During treatment, press to display treatment parameters for channel 3 and 4.
 - Rotate to adjust the output intensity for channel 3.

- 8) Channel Button:
- Press to enter treatment parameters for channel 3 and 4.
 - During treatment, press to display treatment parameters for channel 3 and 4.
 - Rotate to adjust the output intensity for channel 4.
- 9) For the eight parameters selection buttons, see below for details:
- B1: In standby mode, press the B1 button a second time to select the therapeutic waveform; In setting mode, press the B1 button to select CC/CV output mode.
- B2: In standby mode, press the B2 button a second time and hold it about 5 seconds to switch to professional therapy or normal therapy; In professional therapy standby mode, press the B2 button to switch treatment step.
- B3: Press the B3 button to select the parameter Time/Cycle./Ramp.
- B4: Press the B4 button to select the therapeutic program. Press and hold the B4 button again to save settings.
- B5: Press the B5 button to select the parameter Vector/F.M/Burst. B6: Press the B6 button to select the parameter C.F./Freq.
- B7: Press the B7 button to select the parameter Beat H./A.M./Duty.
- B8: Press the B8 button to select the parameter Beat L./P.Dur/Freq.

Remark:

CC—Constant Current output mode

CV — Constant Voltage output mode F.M. — Frequency Modulation

Burst— Burst Frequency

Freq. — Frequency

C.F. — Carrier Frequency

Duty — Duty Cycle for Russian Waveform Beat H. — Sweep High Beat Frequency A.M. — Amplitude Modulation

Beat L. — Sweep Low Beat Frequency P.Dur. — Pulse Duration

Time. — Treatment time

Cycle— Cycle time(Con/Rel)


Ramp— Ramp time

10) Adapter Receptacle

11) ON/OFF Switch

12) Output Connector: connect with connector of cable

4.2 User Interface



- 1) Displays output intensity of channel 1;
Displays total treatment steps of channels 1 and 2.
- 2) Displays output intensity of channel 2; Displays treatment steps of channels 1 and 2.
- 3) Displays output intensity of channel 3;
Displays total treatment steps of channels 3 and 4.
- 4) Displays output intensity of channel 4; Displays treatment steps of channels 3 and 4.
- 5) Displays therapeutic programs (including professional mode) of channels 1 and 2.
- 6) Displays treatment time, cycle time or ramp time of channels 1 and 2.
- 7) Displays 7 therapeutic waveform: IF-4P (IFC-Interferential, Traditional 4 Poles), IF-2P (Premodulated, Traditional 2 Poles IFC), TENS, EMS S (Synchronous), EMS A (Asynchronous), or Russian S (Synchronous), and Russian A (Asynchronous).
- 8) Displays parameters of B7 button: Beat H./A.M./Duty.
- 9) Displays parameters of B5 button: Vector/F.M./Burst.
- 10) Displays parameters of B8 button: Beat L./P.Dur./Freq.
- 11) Displays parameters of B6 button: C.F./Freq.
- 12) Displays treatment time, cycle time or ramp time of channels 3 and 4.
- 13) Displays therapeutic programs (including professional mode) of channels 3 and 4.

5. INSTALLATION

5.1 Before Use

Remove the equipment and all accessories from shipping carton and box. Visually check if there is any damage or missing parts or accessories. If yes, please report to local dealer or retailer where you purchase this unit. Your Quattro™ 2.5 – DQ8450 equipment contains the following accessories.



No.	Replenishment Part #	Part	Quantity
1	ER2535B2	Rubber Electrodes, 2,5" x 3,5"	4 pcs
2	ER2743B2	Rubber Electrodes, 2,75" x 4,3"	4 pcs
3	ES2740Y2	Electrode Sponges, 2,75" x 4"	4 pcs
4	ES3047Y2	Electrode Sponges, 3,0" x 4,75"	4 pcs
5		Self-adhesive Electrodes, 2" x 2"	8 pcs
6		Self-adhesive Electrodes, 2" x 3,5"	8 pcs
7	EW2023BW2	Elastic wrap, 3" x 23,5"	2 pcs
8	EW3047BW2	Elastic wrap 3" x 47"	2 pcs
9	WQ6005PLUG	Lead wires	4 pcs
10	DQ2001MGC	Adapter 100-240V/50-60Hz	1 pc
11	DQ8432CPLUG	Connector of cable	2 pcs
12		User Manual	1 pc

5.2 Connection of the Power Adapter

- Connect the power adapter to the device connector.
- Connect the power adapter to a wall socket.



Caution

- Prior to connecting this device to the power supply, check that the voltage and frequency stated on the rating label match with the available power supply.
- The power adapter is a part of the supply circuit on which the device's safety partly depends upon. The approvals for Quattro™ 2.5 are only valid if used in combination with this type of adapter.

5.3 Switching ON

Switch on the device, using ON/OFF switch [☐]. The device executes a self-test, checking all important functions and enters standby mode about 6~8 seconds later.

5.4 Switching OFF and Disconnecting Power Adapter

- Switch off the device by switching the ON/OFF switch from the [☐] to the [○] position.
- Remove the power adapter from the wall socket.
- Remove the power adapter from the device.

6. OPERATION

6.1 Measures with Regards to Treatment

6.1.1 Electrotherapy Before the Treatment

- Ensure there are no contraindications to treatment.
- Inspect the skin treatment area for any abrasions, inflammation, surface veins, etc.
- Clean the skin treatment area with soap or alcohol (70%).
- Shaving or clipping excessive hair on the skin treatment area can provide optimal treatment.
- Test the heat sensitivity of the treatment area.

6.1.2 Electrode Placement

- Examine the skin for any wounds and clean the skin.
- When using self-adhesive electrodes, remove the electrode from plastic backing.
- Apply the electrodes to the treatment area.
- Ensure that the electrodes are applied securely to the skin.
- Ensure good contact between each electrode and the skin.
- Check the electrode contact regularly during the treatment.
- Examine the skin again after the treatment.
- Choose electrodes to properly fit the anatomy.
- Follow electrode manufacturer's instructions for use.
- To avoid skin irritation due to high current density, do not use electrodes smaller in surface area than 25cm² self-adhesive electrodes.

**Caution**

- Keep electrodes separated during treatment. Electrodes in contact with each other could result in improper stimulation or skin burns.
- Output current density is related to electrode size. Improper application may result in patient injury. If any question arises as to the proper electrode size, consult a licensed practitioner prior to the therapy session.
- Powered muscle stimulators should be used only with the leads and electrodes recommended by the manufacturer.

**6.1.3
Self- Adhesive
Electrodes**

This device is supplied with 8 pieces 2" x 2" and 8 pieces 2" x 3.5" adhesive electrodes. You can select the appropriate sized electrodes according to the treatment area and output current density.

It is recommended that manufacturer's electrodes be used whenever possible to ensure the highest level of performance and contact with the treatment area and to give the most uniform delivery of the prescribed electrotherapy treatment.


Properly dispose of used electrodes upon completion of the therapy session. If you are unsure of your electrode's adhesive properties, it is suggested that you order new replacement electrodes. Replacement electrodes should be re-ordered through or on the advice of your physician to ensure proper quality.

Before applying electrodes, be sure the skin surface over which electrodes are placed is thoroughly cleaned and dried. Make sure the electrodes are placed firmly to the skin and make good contact between the skin and the electrodes. Place the electrodes over the skin; attach them properly, firmly, and evenly.


**Caution**

- 1) Before applying the self-adhesive electrodes, it is recommended to wash and degrease the skin, and then dry it.
- 2) Do not turn on the device when the electrodes are not positioned on the body.
- 3) Never remove the self-adhesive electrodes from the skin while the device is still turned on.
- 4) It is recommended that, at minimum, 2"x 2" self-adhering based, square electrodes are used at the treatment area.

Connecting Lead Wires

	<p>Insert the lead with the red (+) electrode connector into one adhesive electrode. Insert the lead with the black (-) electrode connector into the other electrode. Make certain the lead wires are seated completely into the electrodes, and there are no bare metal parts of the pins exposed.</p>
---	---


Securing Electrodes

	<p>Remove the adhesive electrodes from the protective backing and apply to the treatment area as prescribed. Ensure that the entire electrode surface is in contact with the patient's skin by pressing the electrode firmly into place.</p>
---	--


6.1.4. Rubber Electrodes Rubber Electrode Instructions

If used for delivery of electrotherapy, there are two conductive mediums for you to select from, the first one is to use electrode sponges as conductive mediums, and the other is to use a conductive medium such as an FDA approved Transmission Gel (Note: transmission gel is sold separately). The rubber electrodes should be secured to the treatment area using the nylon wraps that are included with the Quattro™ 2.5.


Connecting Lead Wires

	<p>Insert the lead with the red (+) electrode connector into one rubber electrode. Insert the lead with the black (-) electrode connector into the other rubber electrode. Make certain the lead wires are seated completely into the electrodes, and there are no bare metal parts of the pins exposed.</p>
---	--


Conductive Medium 1

	<p>Insert the rubber electrodes into the electrode sponges that have been moistened with distilled water.</p>
---	---

Conductive Medium 2

	<p>Liberally apply transmission gel to electrode prior to placement on the patient. Please note: purchase the transmission gel with CE mark or that has been cleared by the FDA.</p>
---	--





Securing Electrodes




Use Nylon Wrap to secure each rubber electrode into position on the patient.

6.2 Quick Set-up for Electrical Stimulation

	<p>1. In order to turn on the device, press ON/OFF switch to [●] icon which is located on the side of the device</p>
	<p>2. When you turn the Quattro™ 2.5 on, the device will perform a self-check for about 6~8 seconds, and then the default parameters will display the last treatment mode.</p>
	<p>3. Connect the electrode wires to the cable; note the color of the wires and the color marks on the cable, they should correspond.</p> <p>Caution: If you want to use 4 channels, connect all electrode lead wires to two cables.</p>
	<p>4. Plug the cable(s) to the device connector(s), the left connector corresponds to output channel 1 and channel 2; the right connector corresponds to output channel 3 and channel 4.</p>
	<p>5. Connect the electrodes to electrode lead wires.</p>
	<p>6. Place the electrodes on the patient according to section 6.1.</p>
	<p>7. Press channel button (① or ②) to enter channel 1 and channel 2 parameter setting mode.</p>

 <p>Waveform CC/CV</p>	<p>8. There are 7 therapeutic waveforms for you to select. Press the "Waveform" button to choose the therapeutic waveform. Then rotate the Parameters Control Knob () to select one of the following waveform: IF-4P, IF-2P, EMS S, EMS A, TENS, Russian S and Russian A.</p>
 <p>Program Save</p>	<p>9. There are 50 programs for all waveforms. For detailed preset programmable parameters for each program refer to section 6.3 in this manual. To change parameters refer to each waveforms "Set-up Procedure". Press the "Program" button to choose the therapeutic program. Then rotate the Parameters Control Knob () to select the therapeutic programs in the corresponding waveform.</p>
<p>CC</p>	<p>10. Press the "Waveform/CC/CV" button twice to select "CC" or "CV" control mode.</p>
	<p>11. Adjust the output intensity and start treatment by rotating the output intensity adjustment knobs on the control panel. 0.5mA/step or 0.5V/step. The "STIM" symbol in the LCD indicates there is intensity output.</p>
	<p>12. For safety purposes, the load detection function was designed so when the output intensity surpasses 10.0mA/10.0V, and there are no electrodes placed on the patient's skin or the device is not connected to the electrodes, an alarm buzzer will sound and the intensity value will flash and reset to zero.</p> <p>Caution: In TENS or EMS mode when the pulse duration is less than 80µs, the load detection function will activate when the output intensity surpasses 40.0mA/40.0V.</p>
	<p>13. Press the channel button ( or ) to enter channel 3 and 4 parameter setting mode if necessary. Then set the parameters following steps 8 thru 12 above.</p>

	<p>14. To pause current channels' treatment, press the output intensity knob until desired channels' (either 1 and 2 or 3 and 4) flash. Then press below button.</p>
	<p>15. Pulse el botón (⊖) button to stop treatment of the current channel(s). If you want to stop the other channel(s) treatment, press corresponding channel button first, and then press the (⊖) button. This will only stop the selected flashing channels, not all 4 channels.</p>

**6.3
Using Preset
Programs**

Each therapeutic waveform has 10 programs with default parameters, but you can also set and save the parameters according to the patient's need. The default parameters for each program are referenced below:

IFC-4P

Syptoms	Program	Vector	C.F.	Beat L.	Beat H.	Time	Step 1	Step 2	CC/ CV
Acute Pain	1	45°	4 kHz	80 Hz	150 Hz	15 min.			CC
	2	45°	4 kHz	130 Hz	150 Hz	15 min.			CC
	3	45°	4 kHz	80 Hz	120 Hz	15 min.			CC
Chronic Pain	4	45°	4 Hz	5 Hz	15 Hz	15 min.			CC
Acute Edema	5	45°	4 kHz	80 Hz	130 Hz	15 min.			CC
Pain/Edema Combo Treatment	6	45°	4 kHz			15 min.	15 Hz	80 - 120 Hz	CC
IFC Sensory	7	45°	4 kHz	80 Hz	120 Hz	15 min.			CC
IFC Motor Sensory	8	45°	4 kHz	15 Hz	100 Hz	15 min.			CC
IFC Motor	9	45°	4 kHz	2 Hz	4 Hz	15 min.			CC
Inflammation Retardation	10	45°	4 kHz	80 Hz	120 Hz	15 min.			CC

**Premodulated
IFC-2P**

Syptoms	Program	Vector	C.F.	Beat L.	Beat H.	Time	Step 1	Step 2	CC/ CV
Acute Pain	1	45°	4 kHz	80 Hz	150 Hz	15 min.			CC
	2	45°	4 kHz	130 Hz	150 Hz	15 min.			CC
	3	45°	4 kHz	80 Hz	120 Hz	15 min.			CC
Chronic Pain	4	45°	4 Hz	5 Hz	15 Hz	15 min.			CC
Acute Edema	5	45°	4 kHz	80 Hz	130 Hz	15 min.			CC
Pain/Edema Combo Treatment	6	45°	4 kHz			15 min.	15 Hz	80 - 120 Hz	CC
IFC Sensory	7	45°	4 kHz	80 Hz	120 Hz	15 min.			CC
IFC Motor Sensory	8	45°	4 kHz	15 Hz	100 Hz	15 min.			CC
IFC Motor	9	45°	4 kHz	2 Hz	4 Hz	15 min.			CC
Inflammation Retardation	10	45°	4 kHz	80 Hz	120 Hz	15 min.			CC

TENS

Symptoms	Program	Freq.	Pulse Dur.	A.M.	F.M./Burst	Time	CC/CV
Acute Pain, lg site	1	150 Hz	250 µs	100 %	0	20 min.	CC
Acute Pain, sm site	2	150 Hz	50 µs	100 %	0	20 min.	CC
Chronic Pain, lg site	3	150 Hz	250 µs	100 %	0	20 min.	CC
Chronic Pain, sm site	4	150 Hz	50 µs	100 %	0	20 min.	CC
TENS Sensory	5	150 Hz	80 µs	100 %	0	15 min.	CC
TENS Motor	6	150 Hz	200 µs	100 %	0	15 min.	CC
Arthritis	7	150 Hz	90 µs	100 %	0	15 min.	CC
Sciatic Nerve Pain	8	150 Hz	325 µs	100 %	0	15 min.	CC
Shoulder Pain	9	150 Hz	375 µs	100 %	0	15 min.	CC
Low Back Pain	10	150 Hz	100 µs	100 %	0	15 min.	CC

EMS S/A







Symptoms	Program	Freq.	Pulse Dur.	A.M.	F.M.	Con/Rel	Ramp	Time	CC/CV
EMS S: Muscle Re-Ed, lg muscle group	1	30 Hz	250 µs	100 %	0	4/12	5 s	15 min.	CC
EMS S: Muscle Re-Ed, sm muscle group	2	30 Hz	50 µs	100 %	0	10/10	1 s	15 min.	CC
EMS S: Muscle Spasm Reduction	3	50 Hz	150 µs	100 %	0	10/10	2 s	15 min.	CC
EMS S: Spasticity Control	4	100 Hz	150 µs	100 %	0	10/10	2 s	20 min.	CC
EMS S: Acute Edema	5	100 Hz	50 µs	100 %	0	--	2 s	20 min.	CC
EMS S: Chronic Edema	6	50 Hz	250 µs	100 %	0	10/10	2 s	20 min.	CC
EMS S: U.E. Biphasic (PENs)	7	150 Hz	70 µs	60 %	0	4/12	2 s	15 min.	CC
EMS S: L.E. Biphasic (PENs)	8	150 Hz	60 µs	60 %	0	4/12	5 s	20 min.	CC
EMS - A: R.O.M, lg muscle group	1	50 Hz	250 µs	100 %	0	10/30	5 s	15 min.	CC
EMS - A: R.O.M, sm muscle group	2	50 Hz	50 µs	100 %	0	10/30	1 s	15 min.	CC



Russian S/ A




Symptoms	Program	Freq.	Duty	Con/Rel	Ramp	Time	CC/ CV
RUSS-S: Muscle Re-Ed, lg muscle group	1	30 Hz	50%	4/12	5 s	15 min.	CC
RUSS-S: Muscle Re-Ed, sm muscle group	2	30 Hz	50%	4/12	1 s	15 min.	CC
RUSS-S: Muscle Spasm Reduction	3	50 Hz	50%	10/10	3 s	15 min.	CC
RUSS-S: Spasticity Control	4	100 Hz	50%	10/10	1 s	20 min.	CC
RUSS-S: Atrophy Retardation, lg muscle group	5	50 Hz	50%	10/30	5 s	15 min.	CC
RUSS-S: Atrophy Retardation, sm muscle group	6	50 Hz	50%	4/12	2 s	15 min.	CC
RUSS-S: Muscle Strengthening, lg muscle group	7	100 Hz	50%	4/12	5 s	15 min.	CC
RUSS-S: Muscle Strengthening, sm muscle group	8	50 Hz	50%	4/12	2 s	15 min.	CC
RUSS-S: Increase Circulation	9	100 Hz	50%	10/30	5 s	15 min.	CC
RUSS-S: Increase R.O.M.	1	100 Hz	50%	10/30	5 s	15 min.	CC

6.4 Each Stimulation Set-up Procedure



6.4.1 4-Pole Interferential Stimulation Set-up Procedure




	<p>1. To turn on the device, press the ON/OFF switch to [], icon which is located on the side of the device.</p>
	<p>2. When you turn the Quattro™ 2.5 on, the device will perform a self-check for about 6~8 seconds, and then the default parameters will display the last treatment mode.</p>
	<p>3. Press channel button ( or ) to enter channel 1 and channel 2 parameter setting mode.</p>
<p>Waveform CC/CV</p> 	<p>4. Press the "Waveform" button to choose the therapeutic waveform, and then rotate the Parameters Control Knob () to select "IF-4P" waveform.</p>
 <p>Program Save</p>	<p>5. Press the "Program" button to choose the therapeutic program (P1 – P10). Then rotate the Parameters Control Knob () to select the therapeutic programs from P-1 to P-10 (Normal mode) or from S-1 to S-10 (Professional mode).</p>
 <p>Sleep</p> 	<p>6. There are two modes in the Quattro™ 2.5 device: Normal mode and Professional mode. Switch between the two modes by pressing and holding the "Step" button until you hear a "Beep" sound, the program will change from "P-1" (Normal mode) to "S-1" (Professional mode). In professional mode, each program has 2 treatment steps, press the "Step" button again, and the LCD will display as shown in the below figure to the left. Press the "Step" button to select treatment steps from 1 to 2. The parameters of each step can be set according to the following methods.</p>
<p>CC</p>	<p>7. Press the "Waveform/CC/CV" button twice to select "CC" or "CV" control mode.</p>



<p>Vector F.M. Burst</p> 	<p>8. Press the "Vector" button, then rotate the Parameters Control Knob () to set the vector (manual) parameter from 0° to 90°, 15°/step.</p>
<p>Vector</p> 	<p>9. Press the "Vector" button again to change to auto mode, the LCD displays "0%" as shown in the figure to the left. Rotate the Parameters Control Knob () to set the vector (auto) parameter from 0 % to 100%, 20%/step.</p>
<p>Beat H. A.M. Duty</p> 	<p>10. Press the "Beat H." button, and then rotate the Parameters Control Knob () to set the parameter from (Beat. L) Hz to 150Hz, 1Hz/step.</p>
 <p>Beat L. Dur. P Freq.</p>	<p>11. Press the "Beat L." button, and then rotate the Parameters Control Knob () to set the parameter from 1Hz to (Beat. H)Hz, 1Hz/step. (Note: Beat L. parameter will not exceed Beat H. setting)</p>
<p>Time Cycle Ramp</p> 	<p>12. Press the "Time" button, and then rotate the Parameters Control Knob () to set the treatment time from 1 min to 60min, 1 min/step.</p>
 <p>Program Save</p>	<p>13. Press and hold the "Save" button until you hear a "beep" sound if you want to save all settings. Note: The settings will not be saved if you have not followed the steps above.</p>
	<p>14. Place the electrodes on the patient. You will need two electrodes for each channel, four in total as shown in the example figure to the left.</p>
	<p>15. Adjust the output intensity and start treatment by rotating the output intensity adjustment knobs on the control panel. 0.5mA/step or 0.5V/step. The "STIM" symbol in the LCD indicates there is intensity output.</p>

	<p>16. For safety purposes, the load detection function was designed so when the output intensity surpasses 10.0mA/10.0V, and there are no electrodes placed on the patient's skin or the device has not been connected to electrodes, an alarm buzzer will sound and the intensity value will flash and reset to zero.</p>
	<p>17. Press the channel button ( or ) to enter channel 3 and 4 parameter setting mode if necessary. Then set the parameters following steps 4 thru 16 above.</p>
	<p>18. To pause current channels' treatment, press output intensity knob until desired channels (either 1 and 2 or 3 and 4) begin flashing. Then press below button.</p>
	<p>19. Press the () button to stop treatment of the current channel(s). If you want to stop the other channel(s) treatment, press corresponding channel button first, and then press the () button. This will only stop the selected flashing channels, not all 4 channels.</p>






6.4.2 2-Pole Interferential Stimulation Set-Up Procedure





	<p>1. To turn on the device, press the ON/OFF switch to [], icon which is located on the side of the device.</p>
	<p>2. When you turn the Quattro™ 2.5 on, the device will perform a self-check for about 6~8 seconds, and then the default parameters will display the last treatment mode.</p>
	<p>3. Press channel button ( or ) to enter channel 1 and channel 2 parameter setting mode.</p>
<p>Waveform CC/CV</p> 	<p>4. Press the “Waveform” button to choose the therapeutic waveform, and then rotate the Parameters Control Knob () to select “IF-4P” waveform.</p>
 <p>Program Save</p>	<p>5. Press the “Program” button to choose the therapeutic program (P1 – P10). Then rotate the Parameters Control Knob () to select the therapeutic programs from P-1 to P-10 (Normal mode) or from S-1 to S-10 (Professional mode).</p>
 <p>Sleep</p> 	<p>6. There are two modes in the Quattro™ 2.5 device: Normal mode and Professional mode. Switch between the two modes by pressing and holding the “Step” button until you hear a “Beep” sound, the program will change from “P-1” (Normal mode) to “S-1” (Professional mode). In professional mode, each program has 2 treatment steps, press the “Step” button again, and the LCD will display as shown in the below figure to the left. Press the “Step” button to select treatment steps from 1 to 2. The parameters of each step can be set according to the following methods.</p>
<p>CC</p>	<p>7. Press the “Waveform/CC/CV” button twice to select “CC” or “CV” control mode.</p>





<p>Beat H. A.M. Duty</p> 	<p>8. Press the "Beat H." button, and then rotate the Parameters Control Knob () to set the parameter from (Beat. L) Hz to 150Hz, 1Hz/step.</p>
 <p>Beat L. Dur. P Freq.</p>	<p>9. Press the "Beat L." button, and then rotate the Parameters Control Knob () to set the parameter from 1Hz to (Beat. H) Hz, 1Hz/step. (Note: Beat L parameter will not exceed Beat H. setting)</p>
<p>Time Cycle Ramp</p> 	<p>10. Press the "Time" button, and then rotate the Parameters Control Knob () to set the treatment time from 1min to 60min, 1min/step.</p>
 <p>Program Save</p>	<p>11. Press and hold the "Save" button until you hear a "beep" sound if you want to save all settings. Note: The settings will not be saved if you have not followed the steps above.</p>
	<p>12. Place the electrodes on the patient. Depending on patient need, one or two channels can be used.</p>
	<p>13. Adjust the output intensity and start treatment by rotating the output intensity adjustment knobs on the control panel. 0.5mA/step or 0.5V/step. The "STIM" symbol in the LCD indicates there is intensity output.</p>
	<p>14. For safety purposes, the load detection function was designed so when the output intensity surpasses 10.0mA/10.0V, and there are no electrodes placed on the patient's skin or the device has not been connected to electrodes, an alarm buzzer will sound and the intensity value will flash and reset to zero.</p>


	<p>15. Press the channel button (③ or ④) to enter channel 3 and 4 parameter setting mode if necessary. Then set the parameters following steps 4 thru 14 above.</p>
	<p>16. To pause current channels' treatment, press output intensity knob until desired channels (either 1 and 2 or 3 and 4) begin flashing. Then press below button.</p>
	<p>17. Press the (⏏) button to stop treatment of the current channel(s). If you want to stop the other channel(s) treatment, press corresponding channel button first, and then press the (⏏) button. This will only stop the selected flashing channels, not all 4 channels.</p>

**6.4.3
EMS S/A
Stimulation
Set-up
Procedure**


	<p>1. To turn on the device, press the ON/OFF switch to [⏻], icon which is located on the side of the device.</p>
	<p>2. When you turn the Quattro™ 2.5 on, the device will perform a self-check for about 6~8 seconds, and then the default parameters will display the last treatment mode.</p>
	<p>3. Press channel button (① or ②) to enter channel 1 and channel 2 parameter setting mode.</p>
<p>Waveform CC/CV</p> 	<p>4. Press the "Waveform" button to choose the therapeutic waveform. Then rotate the Parameters Control Knob (⦿) to select "EMS" or "EMS A" waveform.</p>
 <p>Program</p>	<p>5. Press the "Program" button to choose the therapeutic program. Then rotate the Parameters Control Knob (⦿) to select the therapeutic programs from P-1 to P-8(Normal mode in EMS S waveform)or from S-1 to S-8 (Professional mode in EMS S waveform). There are 2 programs (P-01~02 or S-01~02) for you to select in EMS A waveform.</p>

	<p>6. There are two modes in Quattro™ 2.5: Normal mode and Professional mode. Switch between the two modes by pressing and holding the "Step" button until you hear a "Beep" sound, the program will change from "P-1" (Normal mode) to "S-1" (Professional mode). In professional mode, each program has 2 treatment steps, press the "Step" button again, and the LCD will display as shown by the figure to the left. Press the "Step" button to select treatment steps from 1 to 2. The parameters of each step can be set according to the following methods.</p>
<p>CC</p>	<p>7. Press the "Waveform/ CC/CV" button twice to select "CC" or "CV" control mode.</p>
<p>Vector F.M. Burst</p> 	<p>8. Press the "F.M." button, and then rotate the parameters control knob () to set the F.M. parameter from 0Hz to 249Hz, 1Hz/step. But F.M.+Freq. ≤ 250Hz.</p>
 <p>C.F. Freq.</p>	<p>9. Press the "Freq." button (B6), and then rotate the Parameters Control Knob () to set the frequency from 1Hz to 250Hz, 1Hz/step. But Freq. ≤ 250-F.M.</p>
<p>Beat H. A.M. Duty</p> 	<p>10. Press the "A.M." button, then rotate the Parameters Control Knob () to set the parameter from 0% to 100% ("A.M." means amplitude modulation, the setting value indicates percentage of modulation, e.g.: "0%" means no modulation, the intensity is outputting at the setting value continuously. "100%" means the output intensity is modulated between 0 and the setting value. The "A.M. waveform" patterns are shown on page 29, 20%/step.</p>
 <p>Beat L. Dur. P Freq.</p>	<p>11. Press the "P.Dur." button, and then rotate the Parameters Control Knob () to set the pulse duration from 30µs to 400µs, 5µs/step.</p>
<p>Time Cycle Ramp</p> 	<p>12. Press the "time" button, and then rotate the Parameters Control Knob () to set the treatment time from 1min to 60min, 1 min/step.</p>

	<p>13. Press the "Time" button again to choose Cycle time, and then rotate the Parameters Control Knob (⊖) to select the cycle time (Contr./Relax) from "-/-(continuous)", "4/4", "4/8", "7/7", "5/5", "4/12", "10/10", "10/20", "10/30" and "10/50".</p>
	<p>14. Press the "Time" button again to choose Ramp time parameters, and then rotate the Parameters Control Knob (⊖) to select the ramp time from 1s, 2s and 5s.</p>
	<p>15. Press and hold the "Save" button until you hear a "beep" sound if you want to save all settings. Note: The settings will not be saved if you have not followed the steps above.</p>
	<p>16. Place the electrodes on the patient. Depending on patient need, one or two channels can be used.</p>
	<p>17. Adjust the output intensity and start treatment by rotating the output intensity adjustable knobs on the control panel. 0.5mA/step or 0.5V/step. The "STIM" symbol in the LCD indicates there is intensity output.</p>
	<p>18. For safety purposes, the load detection function was designed so when the output intensity surpasses 10.0mA/10.0V, and there are no electrodes placed on the patient's skin or the device has not been connected to electrodes, an alarm buzzer will sound and the intensity value will flash and reset to zero. Note: If the pulse duration is less than 80µs, the load detection function will activate when output intensity surpasses or equal 40.0mA/40.0V.</p>
	<p>19. Press the channel button (Ⓜ₃ or Ⓜ₄) to enter channel 3 and 4 parameter setting mode if necessary. Then set the parameters following steps 4 thru 18 above.</p>

	<p>20. To pause current channels' treatment, press output intensity knob until desired channels (either 1 and 2 or 3 and 4) begin flashing. Then press below button.</p>
	<p>21. Press the () button to stop treatment of the current channel(s). If you want to stop the other channel(s) treatment, press corresponding channel button first, and then press the () button. This will only stop the selected flashing channels, not all 4 channels.</p>





A.M. waveform:











Remark:

- 1) EMS S (Synchronous) stimulation: For this treatment, use two channels (either channels 1 & 2 or 3 & 4). Electrodes for each channel will be placed on different muscle groups. During treatment, both channels will stimulate at the same time. Best used for Bi-lateral conditions.
- 2) EMS A (Asynchronous) stimulation: For this treatment, use two channels (either 1 & 2 or 3 & 4). Electrodes for each channel will be placed on separate muscle groups. During treatment, channels will alternate stimulation. Best used when treating opposing muscle groups.




6.4.4 TENS Stimulation Set-up Procedure




	<p>1. To turn on the device, press the ON/OFF switch to (⊙) icon which is located on the side of the device.</p>
	<p>2. When you turn the Quattro™ 2.5 on, the device will perform a self-check for about 6~8 seconds, and then the default parameters will display the last treatment mode.</p>
	<p>3. Press channel button (① or ②) to enter channel 1 and channel 2 parameter setting mode.</p>
<p>Waveform CC/CV</p> 	<p>4. Press the "Waveform" button to choose the therapeutic waveform. Then rotate the Parameters Control Knob (⊙) to select "TENS" waveform.</p>
 <p>Program Save</p>	<p>5. Press the "Program" button to choose the therapeutic program. Then rotate the Parameters Control Knob (⊙) to select the therapeutic programs from P-1 to P-10 (Normal mode) or from S-1 to S-10 (Professional mode).</p>
 <p>Step</p> 	<p>6. There are two modes in Quattro™ 2.5: Normal mode and Professional mode. Switch between the two modes by pressing and holding the "Step" button until you hear a "Beep" sound, the program will change from "P-1" (Normal mode) to "S-1" (Professional mode). In professional mode, each program has 2 treatment steps, press the "Step" button again, and the LCD will display as shown by the below figure to the left. Press the "Step" button to select treatment steps from 1 to 2. The parameters of each step can be set according to the following methods.</p>
<p>CC</p>	<p>7. Press the "Waveform/CC/CV" button twice to select "CC" or "CV" control mode.</p>



<p>Vector F.M. Burst</p> 	<p>8. Press the "F.M." button, and then rotate the Parameters Control Knob () to set the F.M. parameter from 0Hz to 249Hz, 1Hz/step. But F.M.+Freq. ≤ 250Hz.</p>
<p>Vector</p> 	<p>9. Press the "F.M." button again to switch to Burst, and then rotate the Parameters Control Knob () to set the Burst rate from 0Hz to 10Hz, 1Hz/step. But Burst × 8 ≤ Freq.</p>
<p>C.F. Freq.</p> 	<p>10. Press the "Freq." button, and then rotate the Parameters Control Knob () to set the frequency from 1Hz to 250Hz, 1Hz/step. But Freq. ≥ Burst × 8 or Freq. ≤ 250-F.M.</p>
<p>Beat H. A.M. Duty</p> 	<p>11. Press the "A.M." button, then rotate the Parameters Control Knob () to set the parameter from 0% to 100% ("A.M." means amplitude modulation, the setting value indicates percentage of modulation, e.g.: "0%" means no modulation, the intensity is outputting at the setting value continuously. "100%" means the output intensity is modulated between 0 and the setting value. The "A.M. waveform" patterns shown on page 29), 20%/step.</p>
<p>Beat L. Dur. P Freq.</p> 	<p>12. Press the "P.Dur." button, and then rotate the Parameters Control Knob () to set the pulse duration from 30µs to 400µs, 5µs/step.</p>
<p>Time Cycle Ramp</p> 	<p>13. Press the "Time" button, and then rotate the Parameters Control Knob () to set the treatment time from 1 min to 60min, 1min/step.</p>
<p>Cycle</p> 	<p>14. Press the "Time" button again to switch to Cycle time, and then rotate the Parameters Control Knob () to select the cycle time (Contr./Relax) from "-/- (continuous)", "4/4", "4/8", "7/7", "5/5", "4/12", "10/10", "10/20", "10/30" and "10/50".</p>
<p>Program Save</p> 	<p>15. Press and hold the "Save" button until you hear a "beep" sound if you want to save all settings. Note: The settings will not be saved if you have not followed the step above.</p>

	<p>16. Place the electrodes on the patient. Depending on patient need, one or two channels can be used.</p>
	<p>17. Adjust the output intensity and start treatment by rotating the output intensity adjustable knobs on the control panel. 0.5mA/step or 0.5V/step. The "STIM" symbol in the LCD indicates there is intensity output.</p>
	<p>18. For safety purposes, the load detection function was designed so when the output intensity surpasses 10.0mA/10.0V, and there are no electrodes placed on the patient's skin or the device has not been connected to electrodes, an alarm buzzer will sound and the intensity value will flash and reset to zero.</p> <p>Note: If the pulse duration is less than 80µs, the load detection function will activate when output intensity surpasses or equal 40.0mA/40.0V.</p>
	<p>19. Press the channel button (③ or ④) to enter channel 3 and 4 parameter setting mode if necessary. Then set the parameters following steps 4 thru 18 above.</p>
	<p>20. To pause current channels' treatment, press output intensity knob until desired channels (either 1 and 2 or 3 and 4) begin flashing. Then press below button.</p>
	<p>21. Press the (▽) button to stop treatment of the current channel(s). If you want to stop the other channel(s) treatment, press corresponding channel button first, and then press the (▽) button. This will only stop the selected flashing channels, not all 4 channels.</p>

6.4.5 Russian S/ A Stimulation Set-up Procedure

	<p>1. To turn on the device, press the ON/OFF switch to (⊙) icon which is located on the side of the device</p>
	<p>2. When you turn the Quattro™ 2.5 on, the device will perform a self-check for about 6~8 seconds, and then the default parameters will display the last treatment mode.</p>
	<p>3. Press channel button (① or ②) to enter channel 1 and channel 2 parameter setting mode.</p>
<p>Waveform CC/CV</p> 	<p>4. Press the "Waveform" button to choose the therapeutic waveform. Then rotate the Parameters Control Knob (⊙) to select "Russian S" or "Russian A" waveform.</p>
 <p>Program Save</p>	<p>5. Press the "Program" button to choose the therapeutic program. Then rotate the Parameters Control Knob (⊙) to select the therapeutic programs from P-1 to P-09 (Normal mode in Russian S waveform) or from S-1 to S-09 (Professional mode in Russian S waveform). There are 1 programs (P-01 or S-01) for you to select in Russian A waveform.</p>
 <p>Step</p> 	<p>6. There are two modes in Quattro™ 2.5: Normal mode and Professional mode. Switch between the two modes by pressing and holding the "Step" button until you hear a "Beep" sound, the program will change from "P-1" (Normal mode) to "S-1" (Professional mode). In professional mode, each program has 2 treatment steps, press the "Step" button again, and the LCD will display as shown by the below figure to the left. Press the "Step" button to select treatment steps from 1 to 2. The parameters of each step can be set according to the following methods.</p>
<p>CC</p>	<p>7. Press the "Waveform/ CC/CV" button twice to select "CC" or "CV" control mode.</p>

	<p>8. Press the "Duty" button, and then rotate the Parameters Control Knob () to set the parameter from 10% to 50%, 10%/step.</p>
	<p>9. Press the "Freq." button and then rotate the Parameters Control Knob () to set the frequency from 20Hz to 100Hz, 5Hz/step.</p>
	<p>10. Press the "Time" button, and then rotate the Parameters Control Knob () to set the treatment time from 1 min to 60 min, 1 min/step.</p>
	<p>11. Press the "Time" button again to choose Cycle time, and then rotate the Parameters Control Knob () to select the cycle time (contr/ relax) from "-/-(continuous)", "5/5", "4/12", "10/10", "10/20", "10/30" and "10/50".</p>
	<p>12. Press the "Time" button again to choose Ramp time, and then rotate the Parameters Control Knob () to select the ramp time from 1 s, 2 s and 5 s.</p>
	<p>13. Press and hold the "Save" button until you hear a "beep" sound if you want to save all settings. Note: The settings will not be saved if you have not followed the step above.</p>
	<p>14. Place the electrodes on the patient. Depending on patient need, one or two channels can be used.</p>
	<p>15. Adjust the output intensity and start treatment by rotating the output intensity adjustable knobs on the control panel. 0.5mA/step or 0.5V/step. The "STIM" symbol in the LCD indicates there is intensity output.</p>

	<p>16. For safety purposes, the load detection function was designed so when the output intensity surpasses 10.0mA/10.0V, and there are no electrodes placed on the patient's skin or the device has not been connected to electrodes, an alarm buzzer will sound and the intensity value will flash and reset to zero.</p>
	<p>17. Press the channel button (③ or ④) to enter channel 3 and 4 parameter setting mode if necessary. Then set the parameters following steps 4 thru 16 above.</p>
	<p>18. To pause current channels' treatment, press output intensity knob until desired channels (either 1 and 2 or 3 and 4) begin flashing. Then press below button.</p>
	<p>19. Press the (⊖) button to stop treatment of the current channel(s). If you want to stop the other channel(s) treatment, press corresponding channel button first, and then press the (⊖) button. This will only stop the selected flashing channels, not all 4 channels.</p>

Remarks:

- 1) There is a "beeping" sound which will appear for approximately 20 seconds to alert the user after the treatment has finished. Press any button to cancel the "beeping" sound.
- 2) If you want to restore factory parameter settings, turn the device off then press and hold knobs (1) and (2) at the same time, and then turn on the device by pressing the ON/OFF switch, keep pressing the (1) and (2) knobs and the device will continuously beep until all parameters are restored to the factory setting.

7. MAINTENANCE

7.1 Cleaning of the Device

Switch off the device and disconnect it from the power supply. The apparatus can be cleaned with a damp cloth. Use lukewarm water and a non-abrasive liquid household cleaner (no abrasive, no alcohol content solution). If a more sterile cleaning is needed, use a cloth moistened with an antimicrobial cleaner.



Caution

Do not submerge the device in liquids. Should the unit become accidentally submerged, contact the dealer or Authorized Service center immediately. Do not attempt to use a system that has been submerged in liquid until inspected and tested by a Service Technician Certified by an Authorized Service center. Do not allow liquids to enter the ventilation holes.

7.2 Cleaning the Electrodes

- Apply the protective backing to the tacky side of the electrode before storing.
- It may be helpful to improve repeated application by spreading a few drops of cold water over the adhesive on the electrode and turn the surface up to air dry. Oversaturation of the electrode with water will reduce the adhesive properties.
- The rubber electrodes should be cleaned with lukewarm water. To disinfect the electrodes or to remove stubborn stains of dirt, use a 70% alcohol solution. The alcohol solution may discolor the electrode; however, this does not affect the operation of the electrodes.
- The sponge pads should be washed in warm water, using a household cleaner. After washing they must be rinsed with clear water, thoroughly drained and then dried. Damaged sponge pads should be replaced. Between uses, store the electrodes in the reusable bag and in a cool dry place.



Caution

- The self-adhesive electrodes are intended for single patient use only.
- If the electrodes do not adhere completely to the patient's skin, it may cause a slight shock.
- If irritation occurs, discontinue use and consult your clinician.
- Always use the electrodes with CE mark, or are legally marketed in the United States under an approved 510(K) procedure.

7.3 Cleaning the lead wires and cables

Periodically wipe the lead wires clean with a cloth dampened in a mild soap solution, and then gently wipe them dry. Use of rubbing alcohol on the lead wires will damage the insulation and dramatically shorten their life.

**7.4
Maintenance**

- Maintenance and all repairs should only be carried out by an authorized agency. The manufacturer will not be held responsible for the results of maintenance or repairs completed by any unauthorized person(s).
- Opening of the equipment by unauthorized agencies is not allowed and will terminate any claim to warranty.

**8.
TROUBLE-
SHOOTING
For optimal
use:**

- Replace lead wires annually.
- Please follow the directions on the electrode packaging for the care of electrodes. The life of the electrodes varies, depending on skin conditions, skin preparation, storage and climate. Replace electrodes that no longer stick.
- NOTE: If the following measures fail to alleviate the problem, please call your dealer.

Problem	Possible Causes	Solution
Displays fail to light up	Adapter contact failure	Ensure adapter is connected. Check the following contacts: <ul style="list-style-type: none"> • All contacts are in place. • All contacts are not broken. • Ensure that adapter is connected.
Stimulation is weak	Electrodes 1. Dried out or contaminated 2. Placement	1. Replace. 2. Electrodes must be a minimum of 2 inches apart.
	Lead wires Old/worn/damaged	Replace.
Stimulation stops	Poor electrode contact	Reapply electrodes, secure firmly.
	Damaged or worn electrodes or lead wires	Replace.
Stimulation is uncomfortable	Intensity is too high	Decrease intensity.
	Electrodes are too close together	Reposition the electrodes.
		Electrodes must be a minimum of 2 inches apart.
	Damaged or worn electrodes or lead wires	Replace.
Electrode active area size is too small.	Replace electrodes with ones that have an active area no less than 25.0 cm ² .	
Stimulation is ineffective.	Improper electrode	Reposition electrode
	Unknown	Contact clinician.
"E1" or "E2" displays on LCD	Hardware problem	Restart the device, if the problem is still exist, please contact the manufacturer or distributor
"E3" displays on LCD	Detected the device is over temperature limit	The device will stop treatment automatically, please wait several minutes before using again.
"E4" displays on LCD	Detected the working current is over the limit	
"E5" displays on LCD	Memorizer failure is detected	Restart the device, if the problem is still exist, please contact the manufacturer or distributor

Remark: If there is a failure, a beeping sound will appear until the failure has been corrected or eliminated, or until the button on the panel has been pressed.

9. SPECIFICATIONS

9.1 General Specifications:

Adapter supply voltage:	100 V - 240 V, 50 Hz a 60 Hz, 0,6 A
Adapter output:	15 V— 1,2 A max.
Type of protection against electric shock:	Class II Equipment
Adapter Dimensions:	88 mm (L)*48 mm (W)*29 mm
Dimensions:	250mm(L)*185mm(L)*82mm(H)
Operating Environmental:	Temperature: 10°C(50°F) to 40°C(104°F), Relative humidity: 30%-85%
Storage Environmental:	Temperature: -20°C(-4°F) to 55°C(131°F), Relative humidity: 20%-90%
Maximum Treatment Time:	60 minutes

9.2 Waveform Specifications

4-Pole Interferential Mode

Waveform Type	Bi-phasic square
Mode Selection	CC (Constant Current) or CV (Constant Voltage)
Vector	Auto: 0%-100% Manual: 0° 90°
Carrier Frequency (C.F.)	4,0 kHz
Sweep High Beat Frequency (Beat H.)	(Beat L.) -150 Hz
Sweep Low Beat Frequency (Beat L.)	1-(Beat H.) Hz
Output Intensity	0-50mA (CC, at 1k ohm load) 0-50V (CV, at 1k ohm load)
Treatment time	1-60 minutes

2-Pole Interferential Mode

Waveform Type	Bi-phasic square
Mode Selection	CC (Constant Current) or CV (Constant Voltage)
Carrier Frequency (C.F.)	4,0 kHz
Sweep High Beat Frequency (Beat H.)	(Beat L.) -150 Hz
Sweep Low Beat Frequency (Beat L.)	1-(Beat H.) Hz
Output Intensity	0-50mA(CC, at 1k ohm load) 0-50V(CV, at 1k ohm load)
Treatment time	1-60 minutes

TENS Mode

Waveform Type	Mono-phasic square
Mode Selection	CC (Constant Current) or CV (Constant Voltage)
Frequency	1-250 Hz
Frequency Modulation (F.M.)	0 a 249Hz
Burst rate (Burst)	0-10 Hz (7 pulse)
Phase duration (Dur. P)	30 - 400 s μ
Amplitude Modulation (A.M.)	0% - 100%
Output Intensity	0-100mA(CC, at 1k ohm load) 0-100V(CV, at 1k ohm load)
Cycle time (Cycle)	Continuous, 4/4, 4/8, 7/7, 5/5, 4/12, 10/10, 10/20, 10/30, 10/50
Treatment time	1-60 minutes

EMS S/A mode

Waveform Type	Bi-phasic square
Mode Selection	CC (Constant Current) or CV (Constant Voltage)
Frequency	1-250 Hz
Frequency Modulation (F.M.)	0 - 249Hz
Step duration (P.Dur.)	30-400 µs
Amplitude Modulation (A.M.)	0% - 100%
Output Intensity	0-100mA(CC, at 1k ohm load) 0-100V(CV, at 1k ohm load)
Treatment time	1-60 minutes
Ramp Time	1s, 2s, 5s
Cycle time (Cycle)	Continuous, 4/4, 4/8, 7/7, 5/5, 4/12, 10/10, 10/20, 10/30, 10/50

Russian Mode

Waveform Type	Bi-phasic square
Mode Selection	CC (Constant Current) or CV (Constant Voltage)
Carrier Frequency (C.F.)	2.5 kHz
Burst Frequency (Freq.)	20 - 100 Hz
Output Intensity	0-50mA(CC, at 1k ohm load) 0-50V(CV, at 1k ohm load)
Duty cycle	10%, 20%, 30%, 40% and 50%.
Cycle time	Continuous, 5/5, 4/12, 10/10, 10/20, 10/30, 10/50.
Treatment time	1-60 minutes
Ramp time	1s, 2s, 5s



Caution

This device has been thoroughly tested and inspected to assure proper performance and operation!

10 STORAGE

For a prolonged pause in treatment, store the device with the adapter in a dry room and protect it against heat, sunshine and moisture. Store the machine in a cool, well-ventilated place. Never place any heavy objects on the machine.

**11
DISPOSAL**



Please dispose of the device in accordance with the directive 2002/96/EC – WEEE (Waste Electrical and Electronic Equipment). Contact your local distributor for information regarding disposal of the unit and accessories.

**12
EMC TABLES**

1. The device requires special precautions regarding electromagnetic compatibility (EMC) and needs to be installed and put into service according to the EMC information supplied in this manual.
2. Care must be taken when operating this device adjacent to or stacked with other equipment. Potential electromagnetic or other interference could occur to this device or other equipment. Try to minimize this interference by not using other equipment in conjunction with it.
3. The performance of the device was determined to be essential. This device has been tested and inspected thoroughly to assure proper performance and operation!

Guidance and manufacturer's declaration - electromagnetic emissions		
The Quattro™ 2.5 is intended for use in the electromagnetic environment specified below. The customer or the user of the Quattro™ 2.5 should ensure that it is used in such an environment.		
Emissions Test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The Quattro™2.5 uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The Quattro™2.5 is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations / flicker emissions IEC 61000-3-3	Not applicable	

Guidance and manufacturer's declaration — electromagnetic immunity			
The Quattro™ 2.5 is intended for use in the electromagnetic environment specified below. The customer or the user of the Quattro™ 2.5 should ensure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Descarga electrostática (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	±2 kV for power supply lines	±2 kV for power supply lines	Main power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV to line(s)	±1 kV line (s) to line (s)	Main power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5% UT (>95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles <5% UT (>95% dip in UT) for 5 seconds	<5% UT (>95% dip in UT) for 0.5 cycle 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles <5% UT (>95% dip in UT) for 5 seconds	Main power quality should be that of a typical commercial or hospital environment. If the user of the device requires continued operation during power mains interruptions, it is needed that the device be powered from an uninterruptible power supply.
Power frequency (50/60Hz) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE: UT is the a.c. main voltage prior to application of the test level.			

Guidance and- manufacturer’s declaration. Electromagnetic immunity			
The Quattro™ 2.5 is intended for use in the electromagnetic environment specified below. The customer or the user of the Quattro™ 2.5 should assure that it is used in such an environment.			
Immunity test	IEC 60501 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6	3 Vrms 150 kHz a 80 MHz	3 Vrms	<p>Portable and mobile RF Communications equipment should be used no closer to any part of the device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance</p> $d = \left[\frac{3.5}{V_1} \right] \sqrt{P}$ $d = \left[\frac{3.5}{E_1} \right] \sqrt{P} \quad 80 \text{ MHz a } 800 \text{ MHz}$ $d = \left[\frac{7}{E_1} \right] \sqrt{P} \quad 800 \text{ MHz a } 2,5 \text{ GHz}$
Radiated RF IEC 61000-4-3	3 V/m 80 MHz a 2,5 GHz	3 V/m	<p>Where P is the maximum output power rating of the transmitter in watts (W) according to the Transmitter manufacturer and d is the recommended separation distance in meters (m).</p> <p>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, a should be less than the compliance level in each frequency range. b</p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p> <div style="text-align: right;">  </div>
<p>NOTE 1 At 80 MHz ends 800 MHz. the higher frequency range applies.</p> <p>NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.</p>			

- a. Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the Quattro™ 2.5 is used exceeds the applicable RF compliance level above, should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the Quattro™ 2.5.
- b. Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

Recommended separation distances between portable and mobile RF communications equipment and the Quattro™ 2.5 device			
The Quattro™ 2.5 device is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Quattro™ 2.5 device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Quattro™ 2.5 as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter		
	150 kHz a 80 MHz	80 MHz a 800 MHz	800 MHz a 2,5 GHz
	$d = \left[\frac{3,5}{V1} \right] \sqrt{P}$	$d = \left[\frac{3,5}{E1} \right] \sqrt{P}$	$d = \left[\frac{7}{E1} \right] \sqrt{P}$
0,01	0,117	0,117	0,233
0,1	0,369	0,369	0,738
1	1,167	1,167	2,333
10	3,689	3,689	7,379
100	11,667	11,667	23,333
For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) accordable to the transmitter manufacturer.			
NOTE 1 At 80 MHz and 800 MHz. the separation distance for the higher frequency range applies.			
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

13.
WARRANTY

Please contact your dealer or the device center in case of a claim under the warranty. If you have to send in the device, enclose a copy of your receipt and state the defect.

A. The following warranty terms apply:

- The warranty period for Quattro™ 2.5 products is two years from date of purchase. In case of a warranty claim, the date of purchase has to be proven by means of the sales receipt or invoice..
- Defects in material or workmanship will be removed free of charge within the warranty period.
- Repairs under warranty do not extend the warranty period either for the device or for the replacement parts.

B. The following is excluded under the warranty:

- Any damage which has arisen due to improper treatment, e.g. nonobservance of the user instruction.
- Any damage which is due to repairs or tampering by the customer or unauthorized third parties.
- Damage which has arisen during transport from the manufacturer to the consumer or during transport to the service center.
- Accessories which are subject to normal wear and tear.
- Liability for direct or indirect consequential losses caused by the unit is excluded even if the damage to the unit is accepted as a warranty claim.

**14.
SÍMBOLOS
NORMALIZADOS**



ON/OFF Switch



Power polarity



Type BF Applied Part



Type of protection against electric shock:
Class II Equipment



Refer to Instruction Manual.



Electrical devices are recyclable material and should not be disposed of with household waste after their useful life! Help us to protect the environment and save resources and take this device to the appropriate collection points.



Please contact the organization which is responsible for waste disposal in your area if you have any questions.



Equipment capable of delivering output values in excess of 10 mA r.m.s. or 10V r.m.s. averaged over any period of 5 s



Stop treatment



Start/Pause the treatment



Serial Number



Manufactured for:
Roscoe Medical, Inc.
21973 Commerce Parkway,
Strongsville, Ohio 44149
www.roscoemedical.com
42-DQ8450_E_00 ©2015